

Certificate of calibration

Type: HFP01-10
Serial number: 10209

Calibration date: 04-06-2014
Cable length: 10 m
Sensitivity: 63.75 $\mu\text{V}/(\text{W}/\text{m}^2)$
Remarks:
Performed by: A. Kanij

Connections:

Cable	Wire	
1	White	Signal [+]
1	Green	Signal [-]
1	Black	Shield

Traceability of calibration is to the "guarded hot plate" of National Physical Laboratory (NPL) of the UK. Applicable standards are ISO 8302 (EN 12664:2001) and ASTM C177.