



To whom it may concern;

Calibration Certificate

Manufacturer : EKO Instruments Co., Ltd.

Calibration date : September 24, 2014

Model : MF-180

Serial number : A1640

Description : Heat Flow Sensor

EKO Ref. number : EX14-212

Cable : 10m

We hereby certify that calibration results of above instrument are as follows;

Sensitivity 0.02429 [mV / W·m⁻²]

Internal Resistance 494 [Ω]

I. Sunaga / Manager of Mfg. Dept.

T. Konuma / Inspector