


To whom it may concern;

Calibration Certificate

Manufacturer :	EKO Instruments Co., Ltd.	Calibration date :	September 24, 2014
Model :	MF-180	Serial number :	A1641
Description :	Heat Flow Sensor	EKO Ref. number :	EX14-212
Cable :	10m		

We hereby certify that calibration results of above instrument are as follows;

Sensitivity	0.02334	[mV / W·m ⁻²]
Internal Resistance	490	[Ω]



I. Sunaga / Manager of Mfg. Dept.



T. Konuma / Inspector